

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2937

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN037652.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169421BS	1,4-Dioxane	0.4	0.30	ug/L	75				70	130	