

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037576.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2937
DFTPP Injection Date: 08/12/2025
DFTPP Injection Time: 15:05

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	87.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	16.2 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037578.D	08/12/2025	16:26
SSTDICC0.2	SSTDICC0.2	BN037579.D	08/12/2025	17:03
SSTDICCC0.4	SSTDICCC0.4	BN037580.D	08/12/2025	17:39
SSTDICC0.8	SSTDICC0.8	BN037581.D	08/12/2025	18:16
SSTDICC1.6	SSTDICC1.6	BN037582.D	08/12/2025	18:52
SSTDICC3.2	SSTDICC3.2	BN037583.D	08/12/2025	19:29
SSTDICC5.0	SSTDICC5.0	BN037584.D	08/12/2025	20:05