

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2939

Client ID : SSTD020678

Date Analyzed: 09/09/2025

Lab File ID: BP025692.D

Time Analyzed: 10:31

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	135835	7.761	560280	10.53	391675	14.38
UPPER LIMIT	271670	8.261	1120560	11.025	783350	14.878
LOWER LIMIT	67917.5	7.261	280140	10.025	195838	13.878
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	101599	7.76	407381	10.54	310555	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.