

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2939</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>08/22/2025 10:00</u>
Lab File ID:	<u>VX047484.D</u>	Init. Calib. Date(s):	<u>08/15/2025 08/15/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:40 12:30</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.637	0.547		-14.13	20
Chloromethane	0.573	0.473	0.1	-17.45	20
Vinyl Chloride	0.720	0.631		-12.36	20
Ethyl Acetate	0.532	0.469		-11.84	20
Bromomethane	0.291	0.250		-14.09	20
Chloroethane	0.440	0.402		-8.64	20
Trichlorofluoromethane	1.065	0.924		-13.24	20
1,1,2-Trichlorotrifluoroethane	0.618	0.560		-9.39	20
Tert butyl alcohol	0.061	0.069		13.11	20
1,1-Dichloroethene	0.630	0.555		-11.9	20
Acrolein	0.129	0.133		3.1	20
Acrylonitrile	0.294	0.310		5.44	20
Acetone	0.263	0.282		7.22	20
Carbon Disulfide	1.900	1.609		-15.32	20
Methyl tert-butyl Ether	1.915	2.042		6.63	20
Methyl Acetate	0.682	0.790		15.84	20
Methylene Chloride	0.697	0.691		-0.86	20
trans-1,2-Dichloroethene	0.668	0.628		-5.99	20
Vinyl Acetate	1.772	1.893		6.83	20
1,1-Dichloroethane	1.222	1.211	0.1	-0.9	20
Cyclohexane	1.126	0.960		-14.74	20
2-Butanone	0.344	0.384		11.63	20
Carbon Tetrachloride	0.560	0.491		-12.32	20
2,2-Dichloropropane	0.922	0.894		-3.04	20
cis-1,2-Dichloroethene	0.778	0.776		-0.26	20
Bromochloromethane	0.508	0.605		19.09	20
Chloroform	1.247	1.259		0.96	20
1,1,1-Trichloroethane	1.073	1.002		-6.62	20
Methylcyclohexane	0.639	0.539		-15.65	20
1,1-Dichloropropene	0.512	0.457		-10.74	20
Benzene	1.533	1.461		-4.7	20
1,2-Dichloroethane	0.543	0.547		0.74	20
Trichloroethene	0.393	0.354		-9.92	20
1,2-Dichloropropane	0.384	0.384		0	20
Dibromomethane	0.279	0.283		1.43	20
Bromodichloromethane	0.578	0.591		2.25	20
4-Methyl-2-Pentanone	0.440	0.451		2.5	20
Toluene	0.947	0.910		-3.91	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2939</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>08/22/2025 10:00</u>
Lab File ID:	<u>VX047484.D</u>	Init. Calib. Date(s):	<u>08/15/2025 08/15/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:40 12:30</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
t-1,3-Dichloropropene	0.505	0.537		6.34	20
cis-1,3-Dichloropropene	0.571	0.594		4.03	20
1,1,2-Trichloroethane	0.360	0.370		2.78	20
1,3-Dichloropropane	0.613	0.627		2.28	20
2-Chloroethyl Vinyl ether	0.211	0.258		22.27	20
2-Hexanone	0.304	0.307		0.99	20
Dibromochloromethane	0.428	0.433		1.17	20
1,2-Dibromoethane	0.371	0.377		1.62	20
Tetrachloroethene	0.362	0.321		-11.33	20
Chlorobenzene	1.188	1.118	0.3	-5.89	20
1,1,1,2-Tetrachloroethane	0.402	0.393		-2.24	20
Hexachloroethane	0.591	0.517		-12.52	20
Ethyl Benzene	2.045	1.893		-7.43	20
m/p-Xylenes	0.763	0.718		-5.9	20
o-Xylene	0.734	0.694		-5.45	20
Styrene	1.240	1.233		-0.56	20
Bromoform	0.302	0.306	0.1	1.33	20
Isopropylbenzene	3.913	3.516		-10.15	20
1,1,2,2-Tetrachloroethane	1.149	1.119	0.3	-2.61	20
1,2,3-Trichloropropane	0.917	0.876		-4.47	20
Bromobenzene	0.955	0.900		-5.76	20
n-propylbenzene	4.673	4.221		-9.67	20
2-Chlorotoluene	2.825	2.553		-9.63	20
1,3,5-Trimethylbenzene	3.219	2.974		-7.61	20
4-Chlorotoluene	3.330	3.046		-8.53	20
tert-Butylbenzene	3.212	2.962		-7.78	20
1,2,4-Trimethylbenzene	3.219	3.002		-6.74	20
sec-Butylbenzene	3.981	3.611		-9.29	20
p-Isopropyltoluene	3.369	3.068		-8.93	20
1,3-Dichlorobenzene	1.825	1.671		-8.44	20
1,4-Dichlorobenzene	1.877	1.682		-10.39	20
n-Butylbenzene	3.133	2.895		-7.6	20
1,2-Dichlorobenzene	1.733	1.613		-6.92	20
1,2-Dibromo-3-Chloropropane	0.222	0.213		-4.05	20
1,2,4-Trichlorobenzene	1.161	1.089		-6.2	20
Hexachlorobutadiene	0.413	0.388		-6.05	20
Naphthalene	3.319	3.315		-0.12	20
1,2,3-Trichlorobenzene	1.095	1.040		-5.02	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2939</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>08/22/2025 10:00</u>
Lab File ID:	<u>VX047484.D</u>	Init. Calib. Date(s):	<u>08/15/2025 08/15/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:40 12:30</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
1,2-Dichloroethane-d4	0.700	0.775		10.71	20
Dibromofluoromethane	0.325	0.344		5.85	20
Toluene-d8	1.160	1.183		1.98	20
4-Bromofluorobenzene	0.428	0.453		5.84	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.