



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02
Lab Code: ACE SDG No.: Q2939
Instrument ID: MSVOA_X Calibration Date/Time: 08/25/2025 10:14
Lab File ID: VX047511.D Init. Calib. Date(s): 08/15/2025 08/15/2025
Heated Purge: (Y/N) N Init. Calib. Time(s): 09:40 12:30
GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
2-Chloroethyl Vinyl ether	250.000	246.13	1.546399	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.