

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081525

Review By	John Carlone	Review On	8/18/2025 8:45:53 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:41:47 AM
SubDirectory	VX081525	HP Acquire Method	HP Processing Method 82X081525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135148
Initial Calibration Stds	VP135153,VP135154,VP135155,VP135156,VP135157,VP135158
CCC	VP135152
Internal Standard/PEM	
ICV/I.BLK	VP135159
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047347.D	15 Aug 2025 08:28		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX047348.D	15 Aug 2025 09:18	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX047349.D	15 Aug 2025 09:40	LR-58,81	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX047350.D	15 Aug 2025 10:01		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX047351.D	15 Aug 2025 10:22		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX047352.D	15 Aug 2025 10:43		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX047353.D	15 Aug 2025 11:05		JC/MD	Ok,M
8	IBLK	IBLK	VX047354.D	15 Aug 2025 11:26		JC/MD	Ok
9	VSTDIC001	VSTDIC001	VX047355.D	15 Aug 2025 12:08	Not used	JC/MD	Not Ok
10	VSTDIC001	VSTDIC001	VX047356.D	15 Aug 2025 12:30		JC/MD	Ok,M
11	VSTDICV050	ICVVX081525	VX047357.D	15 Aug 2025 13:11		JC/MD	Ok,M
12	VX0815WBL01	VX0815WBL01	VX047358.D	15 Aug 2025 13:39		JC/MD	Ok
13	VX0815MBL01	VX0815MBL01	VX047359.D	15 Aug 2025 14:00		JC/MD	Ok
14	VX0815WBS01	VX0815WBS01	VX047360.D	15 Aug 2025 14:21		JC/MD	Ok,M
15	VX0815WBSD01	VX0815WBSD01	VX047361.D	15 Aug 2025 14:49		JC/MD	Ok,M
16	Q2880-01	STORAGE-BLANK-SO	VX047362.D	15 Aug 2025 15:10	vial A pH<2	JC/MD	Ok
17	Q2880-02	STORAGE-BLANK-WA	VX047363.D	15 Aug 2025 15:32	vial A pH<2	JC/MD	Ok
18	Q2880-03	STORAGE-BLANK-WA	VX047364.D	15 Aug 2025 15:53	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX081525

Review By	John Carlone	Review On	8/18/2025 8:45:53 AM
Supervise By	Maresh Dadoda	Supervise On	8/19/2025 1:41:47 AM
SubDirectory	VX081525	HP Acquire Method	HP Processing Method 82X081525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135148		
Initial Calibration Stds	VP135153,VP135154,VP135155,VP135156,VP135157,VP135158		
CCC	VP135152		
Internal Standard/PEM			
ICV/I.BLK	VP135159		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2880-04	STORAGE-BLANK-SAM	VX047365.D	15 Aug 2025 16:15	vial A pH<2	JC/MD	Ok
20	Q2886-01	BP-TT182D-TB-202508	VX047366.D	15 Aug 2025 16:36	vial A pH<2	JC/MD	Ok
21	Q2886-02	BP-TT182D-GW-202508	VX047367.D	15 Aug 2025 16:58	vial A pH<2 Surrogate Fail	JC/MD	ReRun
22	Q2837-01	TW-WTS-13	VX047368.D	15 Aug 2025 17:19	vial B pH<2	JC/MD	Ok,M
23	VSTDCCC050	VSTDCCC050EC	VX047369.D	15 Aug 2025 17:41		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX082225

Review By	John Carlone	Review On	8/25/2025 8:50:19 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 3:23:15 PM
SubDirectory	VX082225	HP Acquire Method	HP Processing Method 82X081525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135234
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135235,VP135236

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047483.D	22 Aug 2025 08:25		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047484.D	22 Aug 2025 10:00	pH#Lot#V12668	JC/MD	Ok,M
3	VX0822MBL01	VX0822MBL01	VX047485.D	22 Aug 2025 10:32		JC/MD	Ok
4	VX0822WBL01	VX0822WBL01	VX047486.D	22 Aug 2025 10:53		JC/MD	Ok
5	VX0822WBS01	VX0822WBS01	VX047487.D	22 Aug 2025 11:22		JC/MD	Ok,M
6	VX0822WBSD01	VX0822WBSD01	VX047488.D	22 Aug 2025 11:51		JC/MD	Ok,M
7	Q2903-07DL	MW-12B-081925DL	VX047489.D	22 Aug 2025 12:12	vial B pH<2	JC/MD	Ok
8	Q2903-13	TB	VX047490.D	22 Aug 2025 12:34	vial A pH<2 TB;Surrogate Fail	JC/MD	ReRun
9	Q2927-01	A6121M	VX047491.D	22 Aug 2025 12:55	vial A pH<2	JC/MD	Ok
10	IBLK	IBLK	VX047492.D	22 Aug 2025 13:16		JC/MD	Ok
11	Q2903-02RE	MW-1C-081925RE	VX047493.D	22 Aug 2025 13:37	vial B pH<2 Surrogate Fail	JC/MD	Confirms
12	Q2903-09	DUP-02-081925	VX047494.D	22 Aug 2025 13:58	vial B pH<2	JC/MD	Ok
13	Q2903-10	MW-17A-081925	VX047495.D	22 Aug 2025 14:19	vial A pH<2	JC/MD	Ok
14	Q2903-11	MW-18A-081925	VX047496.D	22 Aug 2025 14:40	vial A pH<2	JC/MD	Ok
15	Q2903-12	MW-16B-081925	VX047497.D	22 Aug 2025 15:01	vial A pH<2	JC/MD	Ok
16	Q2934-01	ENV-SW01	VX047498.D	22 Aug 2025 15:22	vial A pH<2	JC/MD	Ok
17	Q2934-02	ENV-SW02	VX047499.D	22 Aug 2025 15:43	vial A pH<2	JC/MD	Ok
18	Q2934-03	ENV-SW03	VX047500.D	22 Aug 2025 16:04	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX082225

Review By	John Carlone	Review On	8/25/2025 8:50:19 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 3:23:15 PM
SubDirectory	VX082225	HP Acquire Method	HP Processing Method 82X081525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135234		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135235,VP135236		

19	Q2934-04	ENV-SW04	VX047501.D	22 Aug 2025 16:25	vial A pH<2	JC/MD	Ok
20	IBLK	IBLK	VX047502.D	22 Aug 2025 16:46		JC/MD	Ok
21	Q2939-01	STORAGE-BLANK-SO	VX047503.D	22 Aug 2025 17:07	vial A pH<2 Surrogate Fail	JC/MD	ReRun
22	Q2939-02	STORAGE-BLANK-WA	VX047504.D	22 Aug 2025 17:27	vial A pH<2	JC/MD	Ok
23	Q2939-03	STORAGE-BLANK-WA	VX047505.D	22 Aug 2025 17:48	vial A pH<2	JC/MD	Ok
24	Q2939-04	STORAGE-BLANK-SAI	VX047506.D	22 Aug 2025 18:09	vial A pH<2	JC/MD	Ok
25	Q2903-05MS	MW-6B-081925MS	VX047507.D	22 Aug 2025 18:30	vial A pH<2	JC/MD	Ok,M
26	Q2903-06MSD	MW-6B-081925MSD	VX047508.D	22 Aug 2025 18:51	vial A pH<2	JC/MD	Ok,M
27	VSTDCCC050	VSTDCCC050EC	VX047509.D	22 Aug 2025 19:12		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX082525

Review By	John Carlone	Review On	8/26/2025 8:55:04 AM
Supervise By	Maresh Dadoda	Supervise On	8/26/2025 3:13:43 PM
SubDirectory	VX082525	HP Acquire Method	HP Processing Method 82X081525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135277
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135278,VP135279

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX047510.D	25 Aug 2025 08:12		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX047511.D	25 Aug 2025 10:14	pH#Lot#V12668	JC/MD	Ok,M
3	VX0825MBL01	VX0825MBL01	VX047512.D	25 Aug 2025 10:43		JC/MD	Ok
4	VX0825WBL01	VX0825WBL01	VX047513.D	25 Aug 2025 11:04		JC/MD	Ok
5	VX0825WBS01	VX0825WBS01	VX047514.D	25 Aug 2025 12:03		JC/MD	Ok,M
6	VX0825WBSD01	VX0825WBSD01	VX047515.D	25 Aug 2025 12:30		JC/MD	Ok,M
7	IBLK	IBLK	VX047516.D	25 Aug 2025 12:51		JC/MD	Ok
8	Q2939-01	STORAGE-BLANK-SO	VX047517.D	25 Aug 2025 13:12	vial B pH<2	JC/MD	Ok
9	Q2903-13	TB	VX047518.D	25 Aug 2025 13:32	vial B pH<2 TB	JC/MD	Ok
10	Q2954-11	BRIDGE	VX047519.D	25 Aug 2025 15:30	vial A pH<2	JC/MD	Ok
11	VSTDCCC050	VSTDCCC050EC	VX047520.D	25 Aug 2025 16:13		JC/MD	Ok,M

M : Manual Integration