

Surrogate Summary

SDG No.: Q2939

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2939-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	55.2	110		74	125
		Dibromofluoromethane	50	49.9	100		75	124
		Toluene-d8	50	49.4	99		86	113
		4-Bromofluorobenzene	50	55.0	110		77	121
Q2939-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	60.5	121		74	125
		Dibromofluoromethane	50	51.1	102		75	124
		Toluene-d8	50	48.6	97		86	113
		4-Bromofluorobenzene	50	55.7	111		77	121
Q2939-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	55.7	111		74	125
		Dibromofluoromethane	50	48.8	98		75	124
		Toluene-d8	50	47.3	95		86	113
		4-Bromofluorobenzene	50	53.7	107		77	121
Q2939-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	56.5	113		74	125
		Dibromofluoromethane	50	49.8	100		75	124
		Toluene-d8	50	48.1	96		86	113
		4-Bromofluorobenzene	50	54.6	109		77	121
VX0822WBL01	VX0822WBL01	1,2-Dichloroethane-d4	50	57.9	116		74	125
		Dibromofluoromethane	50	50.7	101		75	124
		Toluene-d8	50	49.3	99		86	113
		4-Bromofluorobenzene	50	56.5	113		77	121
VX0822WBS01	VX0822WBS01	1,2-Dichloroethane-d4	50	52.6	105		74	125
		Dibromofluoromethane	50	48.3	97		75	124
		Toluene-d8	50	46.7	93		86	113
		4-Bromofluorobenzene	50	49.8	99		77	121
VX0822WBSD0	VX0822WBSD01	1,2-Dichloroethane-d4	50	49.6	99		74	125
		Dibromofluoromethane	50	47.9	96		75	124
		Toluene-d8	50	46.5	93		86	113
		4-Bromofluorobenzene	50	48.5	97		77	121
VX0825WBL01	VX0825WBL01	1,2-Dichloroethane-d4	50	54.7	109		74	125
		Dibromofluoromethane	50	49.6	99		75	124
		Toluene-d8	50	48.6	97		86	113
		4-Bromofluorobenzene	50	54.0	108		77	121
VX0825WBS01	VX0825WBS01	1,2-Dichloroethane-d4	50	52.9	106		74	125
		Dibromofluoromethane	50	52.7	105		75	124
		Toluene-d8	50	50.9	102		86	113
		4-Bromofluorobenzene	50	54.1	108		77	121