

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q2939

Lab File ID: VX047510.D

BFB Injection Date: 08/25/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:12

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.7
75	30.0 - 60.0% of mass 95	51.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.2
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	74.3
175	5.0 - 9.0% of mass 174	5.5 (7.4) 1
176	95.0 - 101.0% of mass 174	71 (95.6) 1
177	5.0 - 9.0% of mass 176	4.7 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX047511.D	08/25/2025	10:14
VX0825WBL01	VX0825WBL01	VX047513.D	08/25/2025	11:04
VX0825WBS01	VX0825WBS01	VX047514.D	08/25/2025	12:03
STORAGE-BLANK-SOIL-REF-6	Q2939-01	VX047517.D	08/25/2025	13:12