

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG NO.: Q2943
 Lab File ID: VL042880.D Date Analyzed: 08/26/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:29
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	122810	2.79	336296	3.97	294948	8.89
UPPER LIMIT	171934	3.12	470814	4.30	412927	9.22
LOWER LIMIT	73686	2.46	201778	3.64	176969	8.56
EPA SAMPLE NO.						
SV1	127308	2.79	346383	3.96	304552	8.89
SV1DL	123837	2.79	342269	3.96	292713	8.89
SV1DUP	127089	2.79	351085	3.97	304927	8.89
VL0826ABL01	123348	2.79	337393	3.96	287495	8.89
VL0826ABS01	120328	2.79	335593	3.96	291587	8.89

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.