



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB136957

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	08/25/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	08/25/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	08/25/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB136976

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	08/27/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.98	1	98	90-110	08/27/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	08/27/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB137001

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.1	0.099	101	85-115	08/28/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	08/28/2025
Sample ID: Cyanide	CCV2	mg/L	0.26	0.25	104	90-110	08/28/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB137012

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Free Cyanide	mg/L	0.099	0.099	100	90-110	08/29/2025
Sample ID: CCV1 Free Cyanide	mg/L	0.25	0.25	100	90-110	08/29/2025
Sample ID: CCV2 Free Cyanide	mg/L	0.25	0.25	100	90-110	08/29/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB137012

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------