



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB136957

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	08/25/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	08/25/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	08/25/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB136976

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	08/27/2025
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	08/27/2025
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.030	0.1	08/27/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB137001

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	08/28/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	08/28/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	08/28/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB137012

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Free Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	08/29/2025
Sample ID: CCB1 Free Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	08/29/2025
Sample ID: CCB2 Free Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	08/29/2025

Initial and Continuing Calibration Blank Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2950

Project: Linden Generating Station - 402

RunNo.: LB137012

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------