

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2950
Project:	Linden Generating Station - 402	Sample ID:	Q2940-01
Client ID:	COMP(1-6)MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	0.99		0.030	U	1	1	99		08/27/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2950
Project:	Linden Generating Station - 402	Sample ID:	Q2940-01
Client ID:	COMP(1-6)MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	1.00		0.030	U	1	1	100		08/27/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2950
Project:	Linden Generating Station - 402	Sample ID:	Q2950-01
Client ID:	402MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.034		0.0015	J	0.04	1	81		08/28/2025
Hexavalent Chromium	mg/L	90-111	0.97		0.0030	U	1.0	2	97		08/25/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2950
Project:	Linden Generating Station - 402	Sample ID:	Q2950-01
Client ID:	402MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.034		0.0015	J	0.04	1	81		08/28/2025
Hexavalent Chromium	mg/L	90-111	0.98		0.0030	U	1.0	2	98		08/25/2025