

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q2950
Project: Linden Generating Station - 402	Sample ID: Q2940-01
Client ID: COMP(1-6)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	0.030	U	0.030	U	1	0		08/27/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2950
Project:	Linden Generating Station - 402	Sample ID:	Q2940-01
Client ID:	COMP(1-6)MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	0.99		1.00		1	1		08/27/2025

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q2950
Project: Linden Generating Station - 402	Sample ID: Q2950-01
Client ID: 402DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		08/25/2025
Cyanide	mg/L	+/-20	0.0015	J	0.0013	J	1	14		08/28/2025
Free Cyanide	mg/L	+/-20	0.00096	U	0.00096	U	1	0		08/29/2025

Duplicate Sample Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2950
Project:	Linden Generating Station - 402	Sample ID:	Q2950-01
Client ID:	402MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.97		0.98		2	0.62		08/25/2025
Cyanide	mg/L	+/-20	0.034		0.034		1	0		08/28/2025