

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2953
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113262.D **Time Analyzed:** 09:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.744	4.648	4.848	512.910	500.000	2.6
Aroclor-1016-2	4.763	4.666	4.866	525.780	500.000	5.2
Aroclor-1016-3	4.820	4.724	4.924	522.060	500.000	4.4
Aroclor-1016-4	4.938	4.843	5.043	486.900	500.000	-2.6
Aroclor-1016-5	5.195	5.099	5.299	519.280	500.000	3.9
Aroclor-1260-1	6.232	6.136	6.336	535.140	500.000	7.0
Aroclor-1260-2	6.422	6.326	6.526	540.920	500.000	8.2
Aroclor-1260-3	6.787	6.691	6.891	547.160	500.000	9.4
Aroclor-1260-4	7.045	6.951	7.151	550.330	500.000	10.1
Aroclor-1260-5	7.289	7.194	7.394	556.000	500.000	11.2
Decachlorobiphenyl	8.676	8.582	8.782	55.790	50.000	11.6
Tetrachloro-m-xylene	3.660	3.563	3.763	50.900	50.000	1.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2953
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113262.D **Time Analyzed:** 09:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.629	4.829	484.100	500.000	-3.2
Aroclor-1016-2	4.743	4.647	4.847	487.100	500.000	-2.6
Aroclor-1016-3	4.918	4.822	5.022	473.180	500.000	-5.4
Aroclor-1016-4	4.960	4.864	5.064	466.820	500.000	-6.6
Aroclor-1016-5	5.172	5.076	5.276	498.890	500.000	-0.2
Aroclor-1260-1	6.199	6.104	6.304	478.900	500.000	-4.2
Aroclor-1260-2	6.388	6.292	6.492	475.750	500.000	-4.9
Aroclor-1260-3	6.539	6.443	6.643	446.690	500.000	-10.7
Aroclor-1260-4	7.008	6.912	7.112	468.480	500.000	-6.3
Aroclor-1260-5	7.250	7.155	7.355	459.060	500.000	-8.2
Decachlorobiphenyl	8.622	8.528	8.728	52.440	50.000	4.9
Tetrachloro-m-xylene	3.652	3.555	3.755	47.810	50.000	-4.4