

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2953
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113277.D **Time Analyzed:** 15:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.746	4.648	4.848	482.230	500.000	-3.6
Aroclor-1016-2	4.764	4.666	4.866	504.100	500.000	0.8
Aroclor-1016-3	4.821	4.724	4.924	500.150	500.000	0.0
Aroclor-1016-4	4.940	4.843	5.043	464.920	500.000	-7.0
Aroclor-1016-5	5.197	5.099	5.299	473.630	500.000	-5.3
Aroclor-1260-1	6.233	6.136	6.336	458.900	500.000	-8.2
Aroclor-1260-2	6.422	6.326	6.526	462.940	500.000	-7.4
Aroclor-1260-3	6.788	6.691	6.891	457.910	500.000	-8.4
Aroclor-1260-4	7.047	6.951	7.151	442.760	500.000	-11.4
Aroclor-1260-5	7.291	7.194	7.394	453.330	500.000	-9.3
Decachlorobiphenyl	8.677	8.582	8.782	45.610	50.000	-8.8
Tetrachloro-m-xylene	3.661	3.563	3.763	48.700	50.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2953
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/27/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113277.D **Time Analyzed:** 15:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.629	4.829	470.550	500.000	-5.9
Aroclor-1016-2	4.744	4.647	4.847	471.540	500.000	-5.7
Aroclor-1016-3	4.919	4.822	5.022	455.740	500.000	-8.9
Aroclor-1016-4	4.962	4.864	5.064	449.320	500.000	-10.1
Aroclor-1016-5	5.174	5.076	5.276	477.230	500.000	-4.6
Aroclor-1260-1	6.200	6.104	6.304	435.610	500.000	-12.9
Aroclor-1260-2	6.389	6.292	6.492	428.880	500.000	-14.2
Aroclor-1260-3	6.540	6.443	6.643	402.840	500.000	-19.4
Aroclor-1260-4	7.009	6.912	7.112	419.970	500.000	-16.0
Aroclor-1260-5	7.252	7.155	7.355	418.050	500.000	-16.4
Decachlorobiphenyl	8.624	8.528	8.728	48.710	50.000	-2.6
Tetrachloro-m-xylene	3.653	3.555	3.755	46.700	50.000	-6.6