

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125

STD. NAME	STD REF.#
Tune/Reschk	VP135214
Initial Calibration Stds	VP135215,VP135216,VP135217,VP135218,VP135219,VP135220
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP135221
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082625

Review By	John Carlone	Review On	8/27/2025 11:39:51 AM
Supervise By	Semsettin Yesilyurt	Supervise On	8/27/2025 3:18:48 PM
SubDirectory	VN082625	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135303
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135304,VP135305

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087667.D	26 Aug 2025 12:36		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087668.D	26 Aug 2025 13:38	pH#Lot#V12668	JC\MD	Ok,M
3	VN0826MBL01	VN0826MBL01	VN087669.D	26 Aug 2025 14:48		JC\MD	Ok
4	VN0826WBL01	VN0826WBL01	VN087670.D	26 Aug 2025 15:08		JC\MD	Ok
5	VN0826WBS01	VN0826WBS01	VN087671.D	26 Aug 2025 15:29		JC\MD	Ok,M
6	VN0826WBSD01	VN0826WBSD01	VN087672.D	26 Aug 2025 16:02		JC\MD	Ok,M
7	Q2963-01	MW1	VN087673.D	26 Aug 2025 16:24	vial A pH<2	JC\MD	Ok
8	Q2964-02	MW3	VN087674.D	26 Aug 2025 16:45	vial A pH<2 Need 5X	JC\MD	Dilution

M : Manual Integration