

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2963
 Lab File ID: VN087668.D Date Analyzed: 08/26/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:38
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	202424	8.21	396757	9.08	378748	11.85
UPPER LIMIT	404848	8.706	793514	9.583	757496	12.347
LOWER LIMIT	101212	7.706	198379	8.583	189374	11.347
EPA SAMPLE NO.						
MW1	169690	8.21	395130	9.08	362382	11.84
VN0826WBL01	165274	8.21	380815	9.08	356736	11.85
VN0826WBS01	198008	8.21	395232	9.08	367348	11.84
VN0826WBSD01	206976	8.21	396217	9.08	359977	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.