



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q2964

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	106	100	106	65 - 135	P	08/27/2025	11:44	LB136987
	Manganese	21.7	20.0	109	65 - 135	P	08/27/2025	11:44	LB136987
	Sodium	2060	2000	103	65 - 135	P	08/27/2025	11:44	LB136987