

**Metals**

- 7 -

**LABORATORY CONTROL SAMPLE SUMMARY**

**Client:** G Environmental

**SDG No.:** Q2964

**Contract:** GENV01

**Lab Code:** ACE

| Analyte    | Units | True Value | Result | C | %<br>Recovery | Acceptance<br>Limits | M |
|------------|-------|------------|--------|---|---------------|----------------------|---|
| PB169418BS |       |            |        |   |               |                      |   |
| Iron       | ug/L  | 1500       | 1490   |   | 99            | 80 - 120             | P |
| Manganese  | ug/L  | 100        | 102    |   | 102           | 80 - 120             | P |
| Sodium     | ug/L  | 1500       | 1410   |   | 94            | 80 - 120             | P |