

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2964
 Lab File ID: VN087676.D Date Analyzed: 08/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:33
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	200815	8.21	393512	9.08	367312	11.85
UPPER LIMIT	401630	8.706	787024	9.583	734624	12.347
LOWER LIMIT	100408	7.706	196756	8.583	183656	11.347
EPA SAMPLE NO.						
RPXY1R	148628	8.21	339597	9.08	324921	11.85
VN0828MBL01	165959	8.21	389198	9.08	354647	11.85
VN0828MBS01	211485	8.21	408603	9.08	370187	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.