

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2964
 Lab File ID: VN087693.D Date Analyzed: 08/29/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:05
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	191520	8.21	371305	9.08	364439	11.85
UPPER LIMIT	383040	8.706	742610	9.583	728878	12.347
LOWER LIMIT	95760	7.706	185653	8.583	182220	11.347
EPA SAMPLE NO.						
RPXY1RDL	144470	8.21	341905	9.08	313398	11.85
VN0829MBL01	149197	8.21	346944	9.08	324171	11.85
VN0829MBS01	188928	8.21	373365	9.08	351090	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.