

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2964
 Lab File ID: VY023214.D Date Analyzed: 08/28/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	542406	7.71	816314	8.62	700282	11.42
UPPER LIMIT	1084810	8.213	1632630	9.116	1400560	11.92
LOWER LIMIT	271203	7.213	408157	8.116	350141	10.92
EPA SAMPLE NO.						
RBBX4R	536714	7.71	1037130	8.62	1043264	11.41
VY0828SBL01	694700	7.71	1286727	8.62	1122725	11.42
VY0828SBS01	498098	7.71	773362	8.62	674652	11.42
VY0828SBSD01	479921	7.71	762688	8.62	660180	11.41

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.