

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2964</u>
Lab File ID:	<u>VN087676.D</u>	Date Analyzed:	<u>08/28/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>11:33</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	185443	13.771				
UPPER LIMIT	370886	14.271				
LOWER LIMIT	92721.5	13.271				
EPA SAMPLE NO.						
RPXY1R	157594	13.77				
VN0828MBL01	163207	13.77				
VN0828MBS01	186353	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.