

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2964
 Lab File ID: VY023214.D Date Analyzed: 08/28/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	350311	13.347				
UPPER LIMIT	700622	13.847				
LOWER LIMIT	175156	12.847				
EPA SAMPLE NO.						
RBBX4R	449321	13.35				
VY0828SBL01	442680	13.35				
VY0828SBS01	338228	13.35				
VY0828SBSD01	335118	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.