

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG No.: Q2964
 Instrument ID: MSVOA_N Calibration Date/Time: 08/26/2025 13:38
 Lab File ID: VN087668.D Init. Calib. Date(s): 08/21/2025 08/21/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 12:09 14:44
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	51.92	-3.830002	20