

Surrogate Summary

SDG No.: **Q2964**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2964-02	MW3	1,2-Dichloroethane-d4	50	53.9	108		70 (74)	130 (125)
		Dibromofluoromethane	50	50.5	101		70 (75)	130 (124)
		Toluene-d8	50	46.1	92		70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.8	94		70 (77)	130 (121)
Q2964-02DL	MW3DL	1,2-Dichloroethane-d4	50	59.5	119		70 (74)	130 (125)
		Dibromofluoromethane	50	54.2	108		70 (75)	130 (124)
		Toluene-d8	50	50.7	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	58.0	116		70 (77)	130 (121)
VN0826WBL01	VN0826WBL01	1,2-Dichloroethane-d4	50	54.6	109		70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101		70 (75)	130 (124)
		Toluene-d8	50	48.3	97		70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.5	93		70 (77)	130 (121)
VN0826WBS01	VN0826WBS01	1,2-Dichloroethane-d4	50	48.0	96		70 (74)	130 (125)
		Dibromofluoromethane	50	46.5	93		70 (75)	130 (124)
		Toluene-d8	50	46.3	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.3	95		70 (77)	130 (121)
VN0826WBSD0	VN0826WBSD01	1,2-Dichloroethane-d4	50	45.6	91		70 (74)	130 (125)
		Dibromofluoromethane	50	47.8	96		70 (75)	130 (124)
		Toluene-d8	50	46.8	94		70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.0	94		70 (77)	130 (121)
VX0827WBL01	VX0827WBL01	1,2-Dichloroethane-d4	50	53.9	108		70 (74)	130 (125)
		Dibromofluoromethane	50	48.5	97		70 (75)	130 (124)
		Toluene-d8	50	47.5	95		70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.2	108		70 (77)	130 (121)
VX0827WBS01	VX0827WBS01	1,2-Dichloroethane-d4	50	58.5	117		70 (74)	130 (125)
		Dibromofluoromethane	50	54.1	108		70 (75)	130 (124)
		Toluene-d8	50	51.2	102		70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.4	113		70 (77)	130 (121)

