

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
 Lab Code: ACE SDG NO.: Q2964  
 Lab File ID: VN087668.D Date Analyzed: 08/26/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 13:38  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 202424        | 8.21  | 396757        | 9.08  | 378748        | 11.85  |
| UPPER LIMIT    | 404848        | 8.706 | 793514        | 9.583 | 757496        | 12.347 |
| LOWER LIMIT    | 101212        | 7.706 | 198379        | 8.583 | 189374        | 11.347 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| MW3            | 162555        | 8.21  | 371496        | 9.08  | 353184        | 11.85  |
| VN0826WBL01    | 165274        | 8.21  | 380815        | 9.08  | 356736        | 11.85  |
| VN0826WBS01    | 198008        | 8.21  | 395232        | 9.08  | 367348        | 11.84  |
| VN0826WBSD01   | 206976        | 8.21  | 396217        | 9.08  | 359977        | 11.85  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.