

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01  
Lab Code: ACE SDG NO.: Q2964  
Lab File ID: VX047534.D Date Analyzed: 08/27/2025  
Instrument ID: MSVOA\_X Time Analyzed: 12:23  
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 222169        | 5.56  | 378947        | 6.76  | 341098        | 10.05  |
| UPPER LIMIT    | 444338        | 6.056 | 757894        | 7.263 | 682196        | 10.549 |
| LOWER LIMIT    | 111085        | 5.056 | 189474        | 6.263 | 170549        | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| MW3DL          | 187177        | 5.56  | 371185        | 6.76  | 366426        | 10.06  |
| VX0827WBL01    | 194583        | 5.57  | 393301        | 6.77  | 391930        | 10.06  |
| VX0827WBS01    | 218139        | 5.56  | 392410        | 6.76  | 366609        | 10.06  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.