

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2964
 Lab File ID: VX047534.D Date Analyzed: 08/27/2025
 Instrument ID: MSVOA_X Time Analyzed: 12:23
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	165835	12.018				
UPPER LIMIT	331670	12.518				
LOWER LIMIT	82917.5	11.518				
EPA SAMPLE NO.						
MW3DL	181024	12.02				
VX0827WBL01	194932	12.02				
VX0827WBS01	195246	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.