

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q2964

Lab File ID: VN087667.D

BFB Injection Date: 08/26/2025

Instrument ID: MSVOA_N

BFB Injection Time: 12:36

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	26.1
75	30.0 - 60.0% of mass 95	58.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.2 (0.2) 1
174	50.0 - 100.0% of mass 95	67.9
175	5.0 - 9.0% of mass 174	4.9 (7.2) 1
176	95.0 - 101.0% of mass 174	65.2 (96) 1
177	5.0 - 9.0% of mass 176	3.9 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN087668.D	08/26/2025	13:38
VN0826WBL01	VN0826WBL01	VN087670.D	08/26/2025	15:08
VN0826WBS01	VN0826WBS01	VN087671.D	08/26/2025	15:29
VN0826WBSD01	VN0826WBSD01	VN087672.D	08/26/2025	16:02
MW3	Q2964-02	VN087674.D	08/26/2025	16:45