

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID: Q2966	OrderDate: 8/27/2025 1:09:00 PM
Client: Aramark Uniforms	Project: Monthly 2025
Contact: Jarrod Mills	Location: D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2966-01	GRAB	WATER	TPH	1664A	08/27/25 12:22		08/28/25 09:00	08/27/25
Q2966-02	COMP	WATER	BOD5	SM5210 B	08/27/25 12:25		08/28/25 16:30	08/27/25
			TSS	SM2540 D			08/29/25 10:00	



SAMPLE DATA

Report of Analysis

Client:	Aramark Uniforms	Date Collected:	08/27/25 12:22
Project:	Monthly 2025	Date Received:	08/27/25
Client Sample ID:	GRAB	SDG No.:	Q2966
Lab Sample ID:	Q2966-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TPH	27.1		1	0.29	5.00	mg/L		08/28/25 09:00	1664A

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Aramark Uniforms	Date Collected:	08/27/25 12:25
Project:	Monthly 2025	Date Received:	08/27/25
Client Sample ID:	COMP	SDG No.:	Q2966
Lab Sample ID:	Q2966-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
BOD5	388		1	0.20	2.00	mg/L		08/28/25 16:30	SM 5210 B-16
TSS	86.0		1	1.00	4.00	mg/L		08/29/25 10:00	SM 2540 D-20

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY

Preparation Blank Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: TPH	LB136991BL mg/L	< 2.5000	2.5000	U	0.29	5.0	08/28/2025
Sample ID: BOD5	LB137003BL mg/L	< 0.2000	0.2000	U	0.20	2.0	08/28/2025
Sample ID: TSS	LB137011BL mg/L	1	2.0000	J	1	4	08/29/2025

Duplicate Sample Summary

Client: Aramark Uniforms	SDG No.: Q2966
Project: Monthly 2025	Sample ID: LB136991BS
Client ID: LB136991BSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.9		17.2		1	1.76		08/28/2025

Duplicate Sample Summary

Client: Aramark Uniforms Project: Monthly 2025 Client ID: GROUNDWATERDUP	SDG No.: Q2966 Sample ID: Q2971-37 Percent Solids for Spike Sample: 0
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
BOD5	mg/L	+/-20	0.20	U	0.20	U	1	0		08/28/2025
TSS	mg/L	+/-5	1840		1830		1	0.03		08/29/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB136991

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136991BS							
TPH	mg/L	20.0	16.9		84	1	78-114	08/28/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB136991

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136991BSD							
TPH	mg/L	20.0	17.2		86	1	78-114	08/28/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB137003

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137003BS							
BOD5	mg/L	198	197		99	1	84.6-115.4	08/28/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB137011

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137011BS							
TSS	mg/L	550	533		97	1	90-110	08/29/2025



RAW DATA

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB136991
Analysis Date: 08/28/2025
BalanceID: WC SC-5
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: rubina
Extraction Date: 08/28/2025
Extraction IN Time: 08:00
Extraction OUT Time: 08:15
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB136991BL	LB136991BL	WATER	1.3	1000	100	2.7453	2.7453	3.02	2.7454	2.7454	0.0001	0.1
2	LB136991BS	LB136991BS	WATER	1.3	1000	100	3.1603	3.1603	3.01	3.1772	3.1772	0.0169	16.9
3	LB136991BSD	LB136991BSD	WATER	1.3	1000	100	3.1585	3.1585	3.03	3.1757	3.1757	0.0172	17.2
4	Q2954-11	BRIDGE	WATER	1.6	1000	100	3.0807	3.0807	3.05	3.0825	3.0825	0.0018	1.8
5	Q2966-01	GRAB	WATER	1.6	1000	100	3.0182	3.0182	3.04	3.0453	3.0453	0.0271	27.1
6	Q2971-37	GROUNDWATER	WATER	1.3	1000	100	3.0602	3.0602	3.03	3.0627	3.0627	0.0025	2.5

QC Batch# LB136991

Test: TPH

Analysis Date: 08/28/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3204
pH Paper 0-14	M6069
Sodium Sulfate	EP2636
1:1 HCL	WP112782
Silica Gel	W3079
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP112783
LCSWD	5.00 ML	WP112784
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 71 °C Dessicator Time In1 : 09:46

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 09:00

Bal Check Time: 08:05 Out OVEN TEMP1: 71 °C Dessicator Time Out1: 10:25

Out Time1: 09:45

After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 70 °C Dessicator Time In2 : 11:31

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 11:00

Bal Check Time: 12:01 Out OVEN TEMP2: 70 °C Dessicator Time Out2: 12:00

Out Time2: 11:30

WORKLIST(Hardcopy Internal Chain)

UB 136991

WorkList Name : tph q2937 WorkList ID : 191539 Department : Wet-Chemistry Date : 08-28-2025 07:35:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2954-11	BRIDGE	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	D41	08/25/2025	1664A
Q2966-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	D31	08/27/2025	1664A
Q2971-37	GROUNDWATER	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	D31	08/27/2025	1664A

Date/Time 08/28/25 07:40
Raw Sample Received by: JD W C
Raw Sample Relinquished by: JD CSM

Date/Time 08/28/25 13:30
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JD W C

BOD5 LOG

ANALYST: rubin
SUPERVISOR: Iwona

Reviewed By: Iwona
On: 9/2/2025 12:29:52 PM
Inst Id: DO METER
LB: LB137003

QC BATCH ID: LB137003
BOD Water: WP114473
Starch: W3149
Sulfuric acid, 1N: WP112832
POLYSEED: WP114475
GGA: WP114474
Chlorine Strips: W3155
pH Strips: W3215

Analysis Date: 08/28/2025
MANGANOUS SULFATE SOLUTION: W3103
Alkaline Iodide Azide: W3109
Sodium Thiosulfate, 0.025N: W3105
NaOH, 1N: WP113878
IncubatorID: INCUBATOR #3
GuageID: 0511064
Zero DO: WP114418

Lab SampleID	Client ID	Bottle No.	VOL. ML	Initial Reading (ML)	Final Reading (ML)	Difference	Average
WINKLER 1	WINKLER 1	1	300	0.0	9.7	9.7	9.7
WINKLER 2	WINKLER 2	2	300	9.9	19.6	9.7	9.7

Meter Calibration1: 8.85 Zero DO Reading1: 0.12 mg/L (<=0.2 Criteria)
Barometric Pressure1: 760 mmHg DO Meter BOD fluid reading for winkler comparison: 9.97

After Incubation

Meter Calibration2: 8.89 Zero DO Reading2: 0.10 mg/L (<=0.2 Criteria)
Barometric Pressure2: 760 mmHg



QC BATCH ID: LB137003

INCUBATOR TEMP IN(C): 20.1

INCUBATOR TEMP OUT(C): 20.1

TIME IN: 16:30

TIME OUT: 11:00

DATE IN: 08/28/2025

DATE OUT: 09/02/2025

Lab SampleID	Bottle No.	Check CL	Initial PH	Final PH	Temp °C	Sam Vol. (mL)	D.O.1 Initial	D.O.2 Final	Depletion	BOD Result (mg/L)	Avg Result (mg/L)	Comment
LB137003BL	1	No	6.65	N/A	20.90	300	9.96	9.94	0.02	0.02	0.02	
POLYSEED	1					10	9.95	6.43	3.52	0.7	0.71	
POLYSEED	2					15	9.91	4.16	5.75	0.77		
POLYSEED	3					20	9.87	3.37	6.5	0.65		
GGA	1					6	9.85	5.26	4.59	194	196.67	
GGA	2					6	9.85	5.19	4.66	197.5		
GGA	3					6	9.82	5.14	4.68	198.5		
Q2966-02	1	No	5.37	7.22	20.10	5	9.89	8.67	-	0	388	pH Adjusted
Q2966-02	2					10	9.85	8.27	-	0		
Q2966-02	3					20	9.82	7.85	-	0		
Q2966-02	4					30	9.80	5.21	4.59	388		
Q2971-37	1	No	6.05	7.12	20.30	5	9.79	8.82	-	0		pH Adjusted
Q2971-37	2					20	9.75	8.51	-	0		
Q2971-37	3					50	9.68	8.11	-	0		
Q2971-37	4					150	9.32	8.00	-	0		
Q2971-37DUP	1	No	6.05	7.12	20.30	5	9.80	8.79	-	0		pH Adjusted
Q2971-37DUP	2					20	9.76	8.64	-	0		
Q2971-37DUP	3					50	9.68	8.20	-	0		
Q2971-37DUP	4					150	9.30	8.05	-	0		

NOTE: 2ml POLYSEED added to GGA and all the Samples, but not in Blank.

NOTE (For, CBOD5): 0.16 g Nitrification Inhibitor added to GGA and all the Samples, but not in Blank.

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 08/28/2025

Run Number: LB137011

BalanceID: WC SC-5

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 08/28/2025 15:30 TEMP1 OUT: 103 °C 08/28/2025 16:30
 TEMP2 IN: 104 °C 08/28/2025 17:00 TEMP2 OUT: 103 °C 08/28/2025 18:00
 TEMP3 IN: 104 °C 08/29/2025 10:00 TEMP3 OUT: 103 °C 08/29/2025 11:30
 TEMP4 IN: 104 °C 08/29/2025 12:00 TEMP4 OUT: 103 °C 08/29/2025 13:30

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB137011BL	LB137011BL	1.3103	1.3103	100	1.3104	1.3104	1.3104	0.0001	1
2	LB137011BS	LB137011BS	1.4753	1.4754	100	1.5286	1.5287	1.5287	0.0533	533
3	Q2966-02	COMP	1.4980	1.4981	300	1.5238	1.5239	1.5239	0.0258	86
4	Q2971-37	GROUNDWATER	1.4812	1.4812	500	2.3989	2.3989	2.3989	0.9177	1835.4
5	Q2971-37DUP	GROUNDWATERDUP	1.4778	1.4778	500	2.3952	2.3952	2.3952	0.9174	1834.8

A = Sample Volume (ml)
 B = Final Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

UB137011

WorkList Name : TSS Q2971

WorkList ID : 191570

Department : Wet-Chemistry

Date : 08-29-2025 07:54:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2966-02	COMP	Water	TSS	Cool 4 deg C	ARAM01	D31	08/27/2025	SM2540 D
Q2971-37	GROUNDWATER	Water	TSS	Cool 4 deg C	PSEG03	D31	08/27/2025	SM2540 D

Date/Time 08/29/25 08:00

Raw Sample Received by: JB WOC

Raw Sample Relinquished by: JDCSM

Date/Time 08/29/25 14:30

Raw Sample Received by: JDCS

Raw Sample Relinquished by: JB WOC

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB136991

Review By	jignesh	Review On	8/28/2025 9:06:00 AM
Supervise By	rubina	Supervise On	8/28/2025 12:51:50 PM
SubDirectory	LB136991	Test	TPH
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3204,M6069,EP2636,WP112782,W3079,NA,WP112783,WP112784,NA		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136991BL	LB136991BL	MB	08/28/25 09:00		jignesh	OK
2	LB136991BS	LB136991BS	LCS	08/28/25 09:00		jignesh	OK
3	LB136991BSD	LB136991BSD	LCSD	08/28/25 09:00		jignesh	OK
4	Q2954-11	BRIDGE	SAM	08/28/25 09:00		jignesh	OK
5	Q2966-01	GRAB	SAM	08/28/25 09:00		jignesh	OK
6	Q2971-37	GROUNDWATER	SAM	08/28/25 09:00		jignesh	OK

Instrument ID: DO METER

Daily Analysis Runlog For Sequence/QC Batch ID # LB137003

Review By	rubina	Review On	9/2/2025 12:29:45 PM
Supervise By	Iwona	Supervise On	9/2/2025 12:29:52 PM
SubDirectory	LB137003	Test	BOD5
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP114473,W3149,WP112832,W3103,W3109,W3105,WP114475,WP114474,WP113878		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB137003BL	LB137003BL	MB	08/28/25 16:30		rubina	OK
2	LB137003BS	LB137003BS	LCS	08/28/25 16:30		rubina	OK
3	Q2966-02	COMP	SAM	08/28/25 16:30		rubina	OK
4	Q2971-37	GROUNDWATER	SAM	08/28/25 16:30		rubina	OK
5	Q2971-37DUP	GROUNDWATERDUP	DUP	08/28/25 16:30		rubina	OK

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB137011

Review By	jignesh	Review On	8/29/2025 11:23:56 AM
Supervise By	Iwona	Supervise On	8/29/2025 11:40:11 AM
SubDirectory	LB137011	Test	TSS
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB137011BL	LB137011BL	MB	08/29/25 10:00		jignesh	OK
2	LB137011BS	LB137011BS	LCS	08/29/25 10:00		jignesh	OK
3	Q2966-02	COMP	SAM	08/29/25 10:00		jignesh	OK
4	Q2971-37	GROUNDWATER	SAM	08/29/25 10:00		jignesh	OK
5	Q2971-37DUP	GROUNDWATERDUP	DUP	08/29/25 10:00		jignesh	OK



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:
COMPANY: Aramark uniforms
ADDRESS: 740 Frelinghuysen Ave
CITY Newark STATE: NJ ZIP: 07114
ATTENTION: Jarrod mills
PHONE: 973-824-1101 FAX:

PROJECT NAME: Monthly
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data) ☐ Other
☐ EDD FORMAT

1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	Grab	W		✓	8-27-25	1:22	1	✓									
2.	comp	W	✓		8-27-25	1:22	1	✓	✓								
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>8-27-25 1:22</u>	RECEIVED BY: 1. <u>[Signature]</u>	DATE/TIME: <u>8-27-25 1:22</u>	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>21.5</u> °C
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME: <u>8-27-25 1:30</u>	RECEIVED BY: 2. <u>[Signature]</u>	DATE/TIME: <u>8-27-25 1:30</u>	Comments:
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>8-27-25 1:30</u>	RECEIVED BY: 3. <u>[Signature]</u>	DATE/TIME: <u>8-27-25 1:30</u>	Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312