

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB136991

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136991BS							
TPH	mg/L	20.0	16.9		84	1	78-114	08/28/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB136991

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136991BSD							
TPH	mg/L	20.0	17.2		86	1	78-114	08/28/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB137003

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137003BS							
BOD5	mg/L	198	197		99	1	84.6-115.4	08/28/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2966

Project: Monthly 2025

Run No.: LB137011

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137011BS							
TSS	mg/L	550	533		97	1	90-110	08/29/2025