

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2967
Project:	Linden Generating Station - 402	Sample ID:	Q2950-01
Client ID:	402MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.034		0.0015	J	0.04	1	81		08/28/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2967
Project:	Linden Generating Station - 402	Sample ID:	Q2950-01
Client ID:	402MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.034		0.0015	J	0.04	1	81		08/28/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2967
Project:	Linden Generating Station - 402	Sample ID:	Q2967-01
Client ID:	402MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.98		0.0030	U	1.0	2	98		08/27/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2967
Project:	Linden Generating Station - 402	Sample ID:	Q2967-01
Client ID:	402MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/L	90-111	0.99		0.0030	U	1.0	2	99		08/27/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2967
Project:	Linden Generating Station - 402	Sample ID:	Q2971-37
Client ID:	GROUNDWATERMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	1.10		0.030	U	1	1	110		08/29/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q2967
Project:	Linden Generating Station - 402	Sample ID:	Q2971-37
Client ID:	GROUNDWATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	1.00		0.030	U	1	1	100		08/29/2025