

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2967

Project: Linden Generating Station - 402

Run No.: LB136984

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	Ib136984BS							
Hexavalent Chromium	mg/L	0.5	0.52		103	1	90-111	08/27/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2967

Project: Linden Generating Station - 402

Run No.: LB137001

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169433BS							
Cyanide	mg/L	0.1	0.10		100	1	85-115	08/28/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q2967

Project: Linden Generating Station - 402

Run No.: LB137017

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169456BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	08/29/2025