

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2972

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	3890	4000	97	95 - 105	P	08/29/2025	11:15	LB137025

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2972

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.8	12.0	98	80 - 120	P	08/29/2025	11:27	LB137025

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2972

Contract: SAFE04

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	5060	5000	101	90 - 110	P	08/29/2025	12:01	LB137025
CCV02	Lead	5070	5000	101	90 - 110	P	08/29/2025	13:03	LB137025
CCV03	Lead	5060	5000	101	90 - 110	P	08/29/2025	14:04	LB137025
CCV04	Lead	5120	5000	102	90 - 110	P	08/29/2025	14:54	LB137025
CCV05	Lead	5090	5000	102	90 - 110	P	08/29/2025	15:47	LB137025
CCV06	Lead	5070	5000	101	90 - 110	P	08/29/2025	16:47	LB137025
CCV07	Lead	5040	5000	101	90 - 110	P	08/29/2025	17:39	LB137025
CCV08	Lead	5020	5000	100	90 - 110	P	08/29/2025	18:18	LB137025