

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Safety Environmental Co. of NY, Inc.

SDG No.: Q2972

Contract: SAFE04

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	2.47			-12	12	08/29/2025	11:40	LB137025
ICSAB01	Lead	49.5	49.0	101	37	61	08/29/2025	11:48	LB137025
ICSA	Lead	44.4			-12	12	08/29/2025	11:52	LB137025
ICSAB	Lead	36.6	49.0	75	37	61	08/29/2025	11:57	LB137025