



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
Data File : VX047556.D
Acq On : 29 Aug 2025 08:45
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Title : SW846 8260
Last Update : Fri Aug 29 02:31:18 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	21085	PASS
75	95	30	60	51.3	57808	PASS
95	95	100	100	100.0	112779	PASS
96	95	5	9	6.5	7306	PASS
173	174	0.00	2	1.0	836	PASS
174	95	50	100	76.5	86229	PASS
175	174	5	9	7.3	6304	PASS
176	174	95	101	95.5	82309	PASS
177	176	5	9	6.6	5453	PASS