

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q2987
 Lab File ID: VX047557.D Date Analyzed: 08/29/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:48
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	172889	12.018				
UPPER LIMIT	345778	12.518				
LOWER LIMIT	86444.5	11.518				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	281595	12.02				
STORAGE-BLANK-WATER-REF-5	200391	12.02				
STORAGE-BLANK-WATER-REF-4	184482	12.02				
STORAGE-BLANK-SAMPLE-MANAGEMENT	196025	12.02				
VX0829WBL01	214482	12.02				
VX0829WBS01	210365	12.02				
VX0829WBSD01	184959	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.