

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV19	Mercury	3.92	4.0	98	90 - 110	CV	09/03/2025	13:16	LB137054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV55	Mercury	5.17	5.0	104	90 - 110	CV	09/03/2025	13:21	LB137054
CCV56	Mercury	5.03	5.0	101	90 - 110	CV	09/03/2025	13:53	LB137054
CCV57	Mercury	5.10	5.0	102	90 - 110	CV	09/03/2025	14:21	LB137054
CCV58	Mercury	5.10	5.0	102	90 - 110	CV	09/03/2025	14:30	LB137054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7900	8000	99	90 - 110	P	09/02/2025	11:30	LB137041
	Antimony	4120	4000	103	90 - 110	P	09/02/2025	11:30	LB137041
	Arsenic	4020	4000	101	90 - 110	P	09/02/2025	11:30	LB137041
	Barium	7990	8000	100	90 - 110	P	09/02/2025	11:30	LB137041
	Beryllium	200	200	100	90 - 110	P	09/02/2025	11:30	LB137041
	Cadmium	1990	2000	100	90 - 110	P	09/02/2025	11:30	LB137041
	Calcium	19700	20000	98	90 - 110	P	09/02/2025	11:30	LB137041
	Chromium	820	800	102	90 - 110	P	09/02/2025	11:30	LB137041
	Cobalt	1990	2000	100	90 - 110	P	09/02/2025	11:30	LB137041
	Copper	1020	1000	102	90 - 110	P	09/02/2025	11:30	LB137041
	Iron	4090	4000	102	90 - 110	P	09/02/2025	11:30	LB137041
	Lead	3920	4000	98	90 - 110	P	09/02/2025	11:30	LB137041
	Magnesium	19600	20000	98	90 - 110	P	09/02/2025	11:30	LB137041
	Manganese	1950	2000	98	90 - 110	P	09/02/2025	11:30	LB137041
	Nickel	1990	2000	100	90 - 110	P	09/02/2025	11:30	LB137041
	Potassium	20200	20000	101	90 - 110	P	09/02/2025	11:30	LB137041
	Selenium	4070	4000	102	90 - 110	P	09/02/2025	11:30	LB137041
	Silver	1030	1000	103	90 - 110	P	09/02/2025	11:30	LB137041
	Sodium	19700	20000	98	90 - 110	P	09/02/2025	11:30	LB137041
	Thallium	4050	4000	101	90 - 110	P	09/02/2025	11:30	LB137041
	Vanadium	1970	2000	99	90 - 110	P	09/02/2025	11:30	LB137041
	Zinc	2030	2000	102	90 - 110	P	09/02/2025	11:30	LB137041

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	112	100	112	80 - 120	P	09/02/2025	11:34	LB137041
	Antimony	49.3	50.0	99	80 - 120	P	09/02/2025	11:34	LB137041
	Arsenic	22.0	20.0	110	80 - 120	P	09/02/2025	11:34	LB137041
	Barium	99.2	100	99	80 - 120	P	09/02/2025	11:34	LB137041
	Beryllium	6.24	6.0	104	80 - 120	P	09/02/2025	11:34	LB137041
	Cadmium	6.01	6.0	100	80 - 120	P	09/02/2025	11:34	LB137041
	Calcium	2090	2000	104	80 - 120	P	09/02/2025	11:34	LB137041
	Chromium	10.2	10.0	102	80 - 120	P	09/02/2025	11:34	LB137041
	Cobalt	29.8	30.0	99	80 - 120	P	09/02/2025	11:34	LB137041
	Copper	21.7	20.0	109	80 - 120	P	09/02/2025	11:34	LB137041
	Iron	114	100	114	80 - 120	P	09/02/2025	11:34	LB137041
	Lead	12.3	12.0	103	80 - 120	P	09/02/2025	11:34	LB137041
	Magnesium	2210	2000	110	80 - 120	P	09/02/2025	11:34	LB137041
	Manganese	21.8	20.0	109	80 - 120	P	09/02/2025	11:34	LB137041
	Nickel	40.3	40.0	101	80 - 120	P	09/02/2025	11:34	LB137041
	Potassium	2050	2000	102	80 - 120	P	09/02/2025	11:34	LB137041
	Selenium	22.6	20.0	113	80 - 120	P	09/02/2025	11:34	LB137041
	Silver	10.5	10.0	106	80 - 120	P	09/02/2025	11:34	LB137041
	Sodium	1870	2000	94	80 - 120	P	09/02/2025	11:34	LB137041
	Thallium	42.1	40.0	105	80 - 120	P	09/02/2025	11:34	LB137041
	Vanadium	40.6	40.0	101	80 - 120	P	09/02/2025	11:34	LB137041
	Zinc	43.6	40.0	109	80 - 120	P	09/02/2025	11:34	LB137041

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, In

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10400	10000	104	90 - 110	P	09/02/2025	12:20	LB137041
	Antimony	5300	5000	106	90 - 110	P	09/02/2025	12:20	LB137041
	Arsenic	5280	5000	106	90 - 110	P	09/02/2025	12:20	LB137041
	Barium	10200	10000	102	90 - 110	P	09/02/2025	12:20	LB137041
	Beryllium	260	250	104	90 - 110	P	09/02/2025	12:20	LB137041
	Cadmium	2620	2500	105	90 - 110	P	09/02/2025	12:20	LB137041
	Calcium	25900	25000	104	90 - 110	P	09/02/2025	12:20	LB137041
	Chromium	1080	1000	108	90 - 110	P	09/02/2025	12:20	LB137041
	Cobalt	2600	2500	104	90 - 110	P	09/02/2025	12:20	LB137041
	Copper	1320	1250	106	90 - 110	P	09/02/2025	12:20	LB137041
	Iron	5300	5000	106	90 - 110	P	09/02/2025	12:20	LB137041
	Lead	5190	5000	104	90 - 110	P	09/02/2025	12:20	LB137041
	Magnesium	25900	25000	104	90 - 110	P	09/02/2025	12:20	LB137041
	Manganese	2540	2500	102	90 - 110	P	09/02/2025	12:20	LB137041
	Nickel	2610	2500	104	90 - 110	P	09/02/2025	12:20	LB137041
	Potassium	26300	25000	105	90 - 110	P	09/02/2025	12:20	LB137041
	Selenium	5360	5000	107	90 - 110	P	09/02/2025	12:20	LB137041
	Silver	1320	1250	106	90 - 110	P	09/02/2025	12:20	LB137041
	Sodium	25600	25000	102	90 - 110	P	09/02/2025	12:20	LB137041
	Thallium	5290	5000	106	90 - 110	P	09/02/2025	12:20	LB137041
	Vanadium	2600	2500	104	90 - 110	P	09/02/2025	12:20	LB137041
	Zinc	2680	2500	107	90 - 110	P	09/02/2025	12:20	LB137041
CCV02	Aluminum	10200	10000	102	90 - 110	P	09/02/2025	13:19	LB137041
	Antimony	5400	5000	108	90 - 110	P	09/02/2025	13:19	LB137041
	Arsenic	5340	5000	107	90 - 110	P	09/02/2025	13:19	LB137041
	Barium	9840	10000	98	90 - 110	P	09/02/2025	13:19	LB137041
	Beryllium	256	250	102	90 - 110	P	09/02/2025	13:19	LB137041
	Cadmium	2590	2500	104	90 - 110	P	09/02/2025	13:19	LB137041
	Calcium	25000	25000	100	90 - 110	P	09/02/2025	13:19	LB137041
	Chromium	1060	1000	106	90 - 110	P	09/02/2025	13:19	LB137041
	Cobalt	2590	2500	104	90 - 110	P	09/02/2025	13:19	LB137041
	Copper	1320	1250	106	90 - 110	P	09/02/2025	13:19	LB137041
	Iron	5040	5000	101	90 - 110	P	09/02/2025	13:19	LB137041
	Lead	5140	5000	103	90 - 110	P	09/02/2025	13:19	LB137041

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24900	25000	100	90 - 110	P	09/02/2025	13:19	LB137041
	Manganese	2430	2500	97	90 - 110	P	09/02/2025	13:19	LB137041
	Nickel	2590	2500	104	90 - 110	P	09/02/2025	13:19	LB137041
	Potassium	25400	25000	102	90 - 110	P	09/02/2025	13:19	LB137041
	Selenium	5460	5000	109	90 - 110	P	09/02/2025	13:19	LB137041
	Silver	1310	1250	105	90 - 110	P	09/02/2025	13:19	LB137041
	Sodium	23300	25000	93	90 - 110	P	09/02/2025	13:19	LB137041
	Thallium	5430	5000	109	90 - 110	P	09/02/2025	13:19	LB137041
	Vanadium	2540	2500	102	90 - 110	P	09/02/2025	13:19	LB137041
	Zinc	2530	2500	101	90 - 110	P	09/02/2025	13:19	LB137041
CCV03	Aluminum	10300	10000	103	90 - 110	P	09/02/2025	14:08	LB137041
	Antimony	5400	5000	108	90 - 110	P	09/02/2025	14:08	LB137041
	Arsenic	5320	5000	106	90 - 110	P	09/02/2025	14:08	LB137041
	Barium	10100	10000	101	90 - 110	P	09/02/2025	14:08	LB137041
	Beryllium	256	250	102	90 - 110	P	09/02/2025	14:08	LB137041
	Cadmium	2590	2500	104	90 - 110	P	09/02/2025	14:08	LB137041
	Calcium	25300	25000	101	90 - 110	P	09/02/2025	14:08	LB137041
	Chromium	1070	1000	107	90 - 110	P	09/02/2025	14:08	LB137041
	Cobalt	2590	2500	104	90 - 110	P	09/02/2025	14:08	LB137041
	Copper	1320	1250	106	90 - 110	P	09/02/2025	14:08	LB137041
	Iron	5290	5000	106	90 - 110	P	09/02/2025	14:08	LB137041
	Lead	5140	5000	103	90 - 110	P	09/02/2025	14:08	LB137041
	Magnesium	25200	25000	101	90 - 110	P	09/02/2025	14:08	LB137041
	Manganese	2480	2500	99	90 - 110	P	09/02/2025	14:08	LB137041
	Nickel	2590	2500	104	90 - 110	P	09/02/2025	14:08	LB137041
	Potassium	26300	25000	105	90 - 110	P	09/02/2025	14:08	LB137041
	Selenium	5460	5000	109	90 - 110	P	09/02/2025	14:08	LB137041
	Silver	1330	1250	106	90 - 110	P	09/02/2025	14:08	LB137041
	Sodium	25000	25000	100	90 - 110	P	09/02/2025	14:08	LB137041
	Thallium	5370	5000	107	90 - 110	P	09/02/2025	14:08	LB137041
CCV04	Vanadium	2560	2500	103	90 - 110	P	09/02/2025	14:08	LB137041
	Zinc	2650	2500	106	90 - 110	P	09/02/2025	14:08	LB137041
	Aluminum	10300	10000	103	90 - 110	P	09/02/2025	14:59	LB137041
	Antimony	5420	5000	108	90 - 110	P	09/02/2025	14:59	LB137041

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5350	5000	107	90 - 110	P	09/02/2025	14:59	LB137041
	Barium	10100	10000	101	90 - 110	P	09/02/2025	14:59	LB137041
	Beryllium	252	250	101	90 - 110	P	09/02/2025	14:59	LB137041
	Cadmium	2590	2500	104	90 - 110	P	09/02/2025	14:59	LB137041
	Calcium	25400	25000	102	90 - 110	P	09/02/2025	14:59	LB137041
	Chromium	1070	1000	107	90 - 110	P	09/02/2025	14:59	LB137041
	Cobalt	2590	2500	104	90 - 110	P	09/02/2025	14:59	LB137041
	Copper	1320	1250	106	90 - 110	P	09/02/2025	14:59	LB137041
	Iron	5350	5000	107	90 - 110	P	09/02/2025	14:59	LB137041
	Lead	5140	5000	103	90 - 110	P	09/02/2025	14:59	LB137041
	Magnesium	25500	25000	102	90 - 110	P	09/02/2025	14:59	LB137041
	Manganese	2500	2500	100	90 - 110	P	09/02/2025	14:59	LB137041
	Nickel	2600	2500	104	90 - 110	P	09/02/2025	14:59	LB137041
	Potassium	26400	25000	106	90 - 110	P	09/02/2025	14:59	LB137041
	Selenium	5490	5000	110	90 - 110	P	09/02/2025	14:59	LB137041
	Silver	1320	1250	106	90 - 110	P	09/02/2025	14:59	LB137041
	Sodium	25600	25000	102	90 - 110	P	09/02/2025	14:59	LB137041
	Thallium	5470	5000	109	90 - 110	P	09/02/2025	14:59	LB137041
	Vanadium	2580	2500	103	90 - 110	P	09/02/2025	14:59	LB137041
	Zinc	2630	2500	105	90 - 110	P	09/02/2025	14:59	LB137041
CCV05	Aluminum	9690	10000	97	90 - 110	P	09/02/2025	16:03	LB137041
	Antimony	5140	5000	103	90 - 110	P	09/02/2025	16:03	LB137041
	Arsenic	5070	5000	101	90 - 110	P	09/02/2025	16:03	LB137041
	Barium	9490	10000	95	90 - 110	P	09/02/2025	16:03	LB137041
	Beryllium	234	250	93	90 - 110	P	09/02/2025	16:03	LB137041
	Cadmium	2420	2500	97	90 - 110	P	09/02/2025	16:03	LB137041
	Calcium	23800	25000	95	90 - 110	P	09/02/2025	16:03	LB137041
	Chromium	984	1000	98	90 - 110	P	09/02/2025	16:03	LB137041
	Cobalt	2440	2500	97	90 - 110	P	09/02/2025	16:03	LB137041
	Copper	1250	1250	100	90 - 110	P	09/02/2025	16:03	LB137041
	Iron	4860	5000	97	90 - 110	P	09/02/2025	16:03	LB137041
	Lead	4800	5000	96	90 - 110	P	09/02/2025	16:03	LB137041
	Magnesium	23800	25000	95	90 - 110	P	09/02/2025	16:03	LB137041
	Manganese	2340	2500	94	90 - 110	P	09/02/2025	16:03	LB137041

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc.

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2440	2500	98	90 - 110	P	09/02/2025	16:03	LB137041
	Potassium	24900	25000	100	90 - 110	P	09/02/2025	16:03	LB137041
	Selenium	5210	5000	104	90 - 110	P	09/02/2025	16:03	LB137041
	Silver	1230	1250	98	90 - 110	P	09/02/2025	16:03	LB137041
	Sodium	23500	25000	94	90 - 110	P	09/02/2025	16:03	LB137041
	Thallium	4980	5000	100	90 - 110	P	09/02/2025	16:03	LB137041
	Vanadium	2420	2500	97	90 - 110	P	09/02/2025	16:03	LB137041
	Zinc	2470	2500	99	90 - 110	P	09/02/2025	16:03	LB137041
CCV06	Aluminum	9710	10000	97	90 - 110	P	09/02/2025	16:54	LB137041
	Antimony	5220	5000	104	90 - 110	P	09/02/2025	16:54	LB137041
	Arsenic	5140	5000	103	90 - 110	P	09/02/2025	16:54	LB137041
	Barium	9580	10000	96	90 - 110	P	09/02/2025	16:54	LB137041
	Beryllium	237	250	95	90 - 110	P	09/02/2025	16:54	LB137041
	Cadmium	2450	2500	98	90 - 110	P	09/02/2025	16:54	LB137041
	Calcium	23800	25000	95	90 - 110	P	09/02/2025	16:54	LB137041
	Chromium	1010	1000	101	90 - 110	P	09/02/2025	16:54	LB137041
	Cobalt	2460	2500	98	90 - 110	P	09/02/2025	16:54	LB137041
	Copper	1260	1250	101	90 - 110	P	09/02/2025	16:54	LB137041
	Iron	5060	5000	101	90 - 110	P	09/02/2025	16:54	LB137041
	Lead	4860	5000	97	90 - 110	P	09/02/2025	16:54	LB137041
	Magnesium	23700	25000	95	90 - 110	P	09/02/2025	16:54	LB137041
	Manganese	2340	2500	93	90 - 110	P	09/02/2025	16:54	LB137041
	Nickel	2460	2500	98	90 - 110	P	09/02/2025	16:54	LB137041
	Potassium	25800	25000	103	90 - 110	P	09/02/2025	16:54	LB137041
	Selenium	5300	5000	106	90 - 110	P	09/02/2025	16:54	LB137041
	Silver	1270	1250	102	90 - 110	P	09/02/2025	16:54	LB137041
	Sodium	23600	25000	94	90 - 110	P	09/02/2025	16:54	LB137041
	Thallium	5080	5000	102	90 - 110	P	09/02/2025	16:54	LB137041
	Vanadium	2430	2500	97	90 - 110	P	09/02/2025	16:54	LB137041
	Zinc	2570	2500	103	90 - 110	P	09/02/2025	16:54	LB137041
CCV07	Aluminum	9570	10000	96	90 - 110	P	09/02/2025	17:43	LB137041
	Antimony	5080	5000	102	90 - 110	P	09/02/2025	17:43	LB137041
	Arsenic	4990	5000	100	90 - 110	P	09/02/2025	17:43	LB137041
	Barium	9600	10000	96	90 - 110	P	09/02/2025	17:43	LB137041

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, In

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	225	250	90	90 - 110	P	09/02/2025	17:43	LB137041
	Cadmium	2380	2500	95	90 - 110	P	09/02/2025	17:43	LB137041
	Calcium	23500	25000	94	90 - 110	P	09/02/2025	17:43	LB137041
	Chromium	971	1000	97	90 - 110	P	09/02/2025	17:43	LB137041
	Cobalt	2390	2500	96	90 - 110	P	09/02/2025	17:43	LB137041
	Copper	1230	1250	98	90 - 110	P	09/02/2025	17:43	LB137041
	Iron	5110	5000	102	90 - 110	P	09/02/2025	17:43	LB137041
	Lead	4700	5000	94	90 - 110	P	09/02/2025	17:43	LB137041
	Magnesium	23300	25000	93	90 - 110	P	09/02/2025	17:43	LB137041
	Manganese	2330	2500	93	90 - 110	P	09/02/2025	17:43	LB137041
	Nickel	2390	2500	96	90 - 110	P	09/02/2025	17:43	LB137041
	Potassium	25700	25000	103	90 - 110	P	09/02/2025	17:43	LB137041
	Selenium	5150	5000	103	90 - 110	P	09/02/2025	17:43	LB137041
	Silver	1230	1250	98	90 - 110	P	09/02/2025	17:43	LB137041
	Sodium	24700	25000	99	90 - 110	P	09/02/2025	17:43	LB137041
	Thallium	4790	5000	96	90 - 110	P	09/02/2025	17:43	LB137041
	Vanadium	2410	2500	96	90 - 110	P	09/02/2025	17:43	LB137041
	Zinc	2500	2500	100	90 - 110	P	09/02/2025	17:43	LB137041
CCV08	Aluminum	9610	10000	96	90 - 110	P	09/02/2025	18:25	LB137041
	Antimony	5150	5000	103	90 - 110	P	09/02/2025	18:25	LB137041
	Arsenic	5040	5000	101	90 - 110	P	09/02/2025	18:25	LB137041
	Barium	9700	10000	97	90 - 110	P	09/02/2025	18:25	LB137041
	Beryllium	229	250	91	90 - 110	P	09/02/2025	18:25	LB137041
	Cadmium	2400	2500	96	90 - 110	P	09/02/2025	18:25	LB137041
	Calcium	23700	25000	95	90 - 110	P	09/02/2025	18:25	LB137041
	Chromium	977	1000	98	90 - 110	P	09/02/2025	18:25	LB137041
	Cobalt	2420	2500	97	90 - 110	P	09/02/2025	18:25	LB137041
	Copper	1240	1250	99	90 - 110	P	09/02/2025	18:25	LB137041
	Iron	5030	5000	101	90 - 110	P	09/02/2025	18:25	LB137041
	Lead	4760	5000	95	90 - 110	P	09/02/2025	18:25	LB137041
	Magnesium	23300	25000	93	90 - 110	P	09/02/2025	18:25	LB137041
	Manganese	2360	2500	94	90 - 110	P	09/02/2025	18:25	LB137041
	Nickel	2410	2500	97	90 - 110	P	09/02/2025	18:25	LB137041
	Potassium	25200	25000	101	90 - 110	P	09/02/2025	18:25	LB137041

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Core Environmental Consultants and Services, Inc

SDG No.: Q2994

Contract: CORE02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5190	5000	104	90 - 110	P	09/02/2025	18:25	LB137041
	Silver	1230	1250	98	90 - 110	P	09/02/2025	18:25	LB137041
	Sodium	24000	25000	96	90 - 110	P	09/02/2025	18:25	LB137041
	Thallium	4760	5000	95	90 - 110	P	09/02/2025	18:25	LB137041
	Vanadium	2420	2500	97	90 - 110	P	09/02/2025	18:25	LB137041
	Zinc	2490	2500	100	90 - 110	P	09/02/2025	18:25	LB137041