

Continuing Calibration Report for SequenceID : FE090325AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE055638.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 7526897.000  | 60.000  | 3.314  | 6.948  | 125448.283 | 128403.808 | 2.302  |
| Aliphatic C12-C16 | 5379583.000  | 40.000  | 6.949  | 10.401 | 134489.575 | 135095.243 | 0.448  |
| Aliphatic C16-C21 | 8682377.000  | 60.000  | 10.402 | 13.781 | 144706.283 | 144217.196 | -0.339 |
| Aliphatic C21-C28 | 11698379.000 | 80.000  | 13.782 | 17.452 | 146229.738 | 143235.734 | -2.090 |
| Aliphatic C28-C40 | 16925194.000 | 120.000 | 17.453 | 22.470 | 141043.283 | 137719.667 | -2.413 |
| Aliphatic EPH     | 50212430.000 | 360.000 | 3.314  | 22.470 | 139478.972 | 138184.135 | -0.937 |