

Continuing Calibration Report for SequenceID : fF090225al

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016298.D

Aliphatic C9-C12	7594242.000	60.000	3.221	6.863	126570.700	118465.299	-6.842
Aliphatic C12-C16	5249179.000	40.000	6.864	10.308	131229.475	122693.653	-6.957
Aliphatic C16-C21	8146821.000	60.000	10.304	13.681	135780.350	128123.078	-5.976
Aliphatic C21-C28	10722331.000	80.000	13.682	17.352	134029.138	125602.733	-6.709
Aliphatic C28-C40	12554888.000	120.000	17.353	22.315	104624.067	101676.216	-2.899
Aliphatic EPH	44267461.000	360.000	3.221	22.315	122965.169	116534.481	-5.518