

Continuing Calibration Report for SequenceID : fF090225al

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016310.D

Aliphatic C9-C12	8001465.000	60.000	3.221	6.863	133357.750	118465.299	-12.571
Aliphatic C12-C16	5545827.000	40.000	6.864	10.308	138645.675	122693.653	-13.002
Aliphatic C16-C21	8616371.000	60.000	10.304	13.681	143606.183	128123.078	-12.085
Aliphatic C21-C28	11411717.000	80.000	13.682	17.352	142646.463	125602.733	-13.570
Aliphatic C28-C40	13694562.000	120.000	17.353	22.315	114121.350	101676.216	-12.240
Aliphatic EPH	47269942.000	360.000	3.221	22.315	131305.394	116534.481	-12.675