

Continuing Calibration Report for SequenceID : fg090225al

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016497.D

Aliphatic C9-C12	7775936.000	60.000	3.289	6.931	129598.933	114749.036	-12.941
Aliphatic C12-C16	5309679.000	40.000	6.932	10.390	132741.975	117373.513	-13.094
Aliphatic C16-C21	7959932.000	60.000	10.391	13.780	132665.533	117474.265	-12.932
Aliphatic C21-C28	10027641.000	80.000	13.781	17.465	125345.513	111788.212	-12.128
Aliphatic C28-C40	10884921.000	120.000	17.466	22.506	90707.675	90884.371	0.194
Aliphatic EPH	41958109.000	360.000	3.289	22.506	116550.303	106882.000	-9.046