

Continuing Calibration Report for SequenceID : fg090225al

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016509.D

Aliphatic C9-C12	7967193.000	60.000	3.289	6.931	132786.550	114749.036	-15.719
Aliphatic C12-C16	5473726.000	40.000	6.932	10.390	136843.150	117373.513	-16.588
Aliphatic C16-C21	8300359.000	60.000	10.391	13.780	138339.317	117474.265	-17.761
Aliphatic C21-C28	10630887.000	80.000	13.781	17.465	132886.088	111788.212	-18.873
Aliphatic C28-C40	12512758.000	120.000	17.466	22.506	104272.983	90884.371	-14.731
Aliphatic EPH	44884923.000	360.000	3.289	22.506	124680.342	106882.000	-16.652