

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q2994  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 09/02/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074770.D **Time Analyzed:** 09:33

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.809  | 5.706     | 5.906  | 490.520            | 500.000           | -1.9 |
| Aroclor-1016-2       | 5.830  | 5.727     | 5.927  | 512.950            | 500.000           | 2.6  |
| Aroclor-1016-3       | 5.893  | 5.790     | 5.990  | 466.300            | 500.000           | -6.7 |
| Aroclor-1016-4       | 5.990  | 5.888     | 6.088  | 492.910            | 500.000           | -1.4 |
| Aroclor-1016-5       | 6.282  | 6.180     | 6.380  | 514.580            | 500.000           | 2.9  |
| Aroclor-1260-1       | 7.399  | 7.297     | 7.497  | 492.680            | 500.000           | -1.5 |
| Aroclor-1260-2       | 7.651  | 7.550     | 7.750  | 486.020            | 500.000           | -2.8 |
| Aroclor-1260-3       | 8.009  | 7.908     | 8.108  | 516.240            | 500.000           | 3.2  |
| Aroclor-1260-4       | 8.237  | 8.136     | 8.336  | 498.280            | 500.000           | -0.3 |
| Aroclor-1260-5       | 8.563  | 8.462     | 8.662  | 488.000            | 500.000           | -2.4 |
| Decachlorobiphenyl   | 10.433 | 10.333    | 10.533 | 49.730             | 50.000            | -0.5 |
| Tetrachloro-m-xylene | 4.657  | 4.554     | 4.754  | 50.280             | 50.000            | 0.6  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CORE02  
**Lab Code:** ACE **SDG NO.:** Q2994  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL01 **Date Analyzed:** 09/02/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074770.D **Time Analyzed:** 09:33

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.898 | 4.799     | 4.999 | 517.610            | 500.000           | 3.5   |
| Aroclor-1016-2       | 4.956 | 4.855     | 5.055 | 532.400            | 500.000           | 6.5   |
| Aroclor-1016-3       | 5.076 | 4.976     | 5.176 | 504.660            | 500.000           | 0.9   |
| Aroclor-1016-4       | 5.117 | 5.017     | 5.217 | 505.230            | 500.000           | 1.0   |
| Aroclor-1016-5       | 5.331 | 5.231     | 5.431 | 479.050            | 500.000           | -4.2  |
| Aroclor-1260-1       | 6.360 | 6.261     | 6.461 | 483.830            | 500.000           | -3.2  |
| Aroclor-1260-2       | 6.546 | 6.448     | 6.648 | 482.450            | 500.000           | -3.5  |
| Aroclor-1260-3       | 6.700 | 6.602     | 6.802 | 449.120            | 500.000           | -10.2 |
| Aroclor-1260-4       | 7.169 | 7.072     | 7.272 | 485.680            | 500.000           | -2.9  |
| Aroclor-1260-5       | 7.408 | 7.311     | 7.511 | 476.460            | 500.000           | -4.7  |
| Decachlorobiphenyl   | 8.811 | 8.715     | 8.915 | 46.670             | 50.000            | -6.7  |
| Tetrachloro-m-xylene | 3.800 | 3.699     | 3.899 | 55.440             | 50.000            | 10.9  |