

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2994
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074785.D **Time Analyzed:** 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	503.890	500.000	0.8
Aroclor-1016-2	5.832	5.727	5.927	516.550	500.000	3.3
Aroclor-1016-3	5.895	5.790	5.990	475.060	500.000	-5.0
Aroclor-1016-4	5.992	5.888	6.088	500.240	500.000	0.0
Aroclor-1016-5	6.285	6.180	6.380	523.750	500.000	4.8
Aroclor-1260-1	7.402	7.297	7.497	503.400	500.000	0.7
Aroclor-1260-2	7.654	7.550	7.750	496.250	500.000	-0.8
Aroclor-1260-3	8.012	7.908	8.108	525.810	500.000	5.2
Aroclor-1260-4	8.240	8.136	8.336	506.790	500.000	1.4
Aroclor-1260-5	8.566	8.462	8.662	497.940	500.000	-0.4
Decachlorobiphenyl	10.436	10.333	10.533	51.080	50.000	2.2
Tetrachloro-m-xylene	4.659	4.554	4.754	50.980	50.000	2.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2994
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074785.D **Time Analyzed:** 16:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.900	4.799	4.999	523.530	500.000	4.7
Aroclor-1016-2	4.957	4.855	5.055	548.400	500.000	9.7
Aroclor-1016-3	5.078	4.976	5.176	522.200	500.000	4.4
Aroclor-1016-4	5.118	5.017	5.217	526.570	500.000	5.3
Aroclor-1016-5	5.332	5.231	5.431	484.650	500.000	-3.1
Aroclor-1260-1	6.363	6.261	6.461	496.290	500.000	-0.7
Aroclor-1260-2	6.548	6.448	6.648	492.000	500.000	-1.6
Aroclor-1260-3	6.702	6.602	6.802	447.940	500.000	-10.4
Aroclor-1260-4	7.172	7.072	7.272	487.340	500.000	-2.5
Aroclor-1260-5	7.411	7.311	7.511	486.670	500.000	-2.7
Decachlorobiphenyl	8.812	8.715	8.915	47.160	50.000	-5.7
Tetrachloro-m-xylene	3.801	3.699	3.899	56.980	50.000	14.0