

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2994
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074791.D **Time Analyzed:** 04:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.820	5.706	5.906	498.190	500.000	-0.4
Aroclor-1016-2	5.840	5.727	5.927	460.870	500.000	-7.8
Aroclor-1016-3	5.903	5.790	5.990	496.800	500.000	-0.6
Aroclor-1016-4	6.000	5.888	6.088	552.460	500.000	10.5
Aroclor-1016-5	6.293	6.180	6.380	508.650	500.000	1.7
Aroclor-1260-1	7.412	7.297	7.497	466.590	500.000	-6.7
Aroclor-1260-2	7.665	7.550	7.750	481.630	500.000	-3.7
Aroclor-1260-3	8.023	7.908	8.108	481.600	500.000	-3.7
Aroclor-1260-4	8.250	8.136	8.336	466.810	500.000	-6.6
Aroclor-1260-5	8.579	8.462	8.662	467.730	500.000	-6.5
Decachlorobiphenyl	10.457	10.333	10.533	47.330	50.000	-5.3
Tetrachloro-m-xylene	4.667	4.554	4.754	52.380	50.000	4.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2994
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074791.D **Time Analyzed:** 04:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.895	4.799	4.999	522.330	500.000	4.5
Aroclor-1016-2	4.953	4.855	5.055	612.200	500.000	22.4
Aroclor-1016-3	5.074	4.976	5.176	570.790	500.000	14.2
Aroclor-1016-4	5.114	5.017	5.217	557.540	500.000	11.5
Aroclor-1016-5	5.329	5.231	5.431	516.170	500.000	3.2
Aroclor-1260-1	6.360	6.261	6.461	477.420	500.000	-4.5
Aroclor-1260-2	6.547	6.448	6.648	473.110	500.000	-5.4
Aroclor-1260-3	6.701	6.602	6.802	420.150	500.000	-16.0
Aroclor-1260-4	7.170	7.072	7.272	429.990	500.000	-14.0
Aroclor-1260-5	7.410	7.311	7.511	421.490	500.000	-15.7
Decachlorobiphenyl	8.813	8.715	8.915	38.910	50.000	-22.2
Tetrachloro-m-xylene	3.795	3.699	3.899	58.700	50.000	17.4