

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2994
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074804.D **Time Analyzed:** 11:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	496.810	500.000	-0.6
Aroclor-1016-2	5.832	5.727	5.927	517.700	500.000	3.5
Aroclor-1016-3	5.895	5.790	5.990	471.190	500.000	-5.8
Aroclor-1016-4	5.992	5.888	6.088	496.800	500.000	-0.6
Aroclor-1016-5	6.285	6.180	6.380	508.400	500.000	1.7
Aroclor-1260-1	7.403	7.297	7.497	460.190	500.000	-8.0
Aroclor-1260-2	7.655	7.550	7.750	448.020	500.000	-10.4
Aroclor-1260-3	8.013	7.908	8.108	464.340	500.000	-7.1
Aroclor-1260-4	8.242	8.136	8.336	446.910	500.000	-10.6
Aroclor-1260-5	8.568	8.462	8.662	441.390	500.000	-11.7
Decachlorobiphenyl	10.441	10.333	10.533	45.450	50.000	-9.1
Tetrachloro-m-xylene	4.659	4.554	4.754	52.030	50.000	4.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CORE02
Lab Code: ACE **SDG NO.:** Q2994
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074804.D **Time Analyzed:** 11:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.799	4.999	510.280	500.000	2.1
Aroclor-1016-2	4.958	4.855	5.055	515.500	500.000	3.1
Aroclor-1016-3	5.076	4.976	5.176	502.690	500.000	0.5
Aroclor-1016-4	5.118	5.017	5.217	500.870	500.000	0.2
Aroclor-1016-5	5.331	5.231	5.431	465.370	500.000	-6.9
Aroclor-1260-1	6.361	6.261	6.461	464.110	500.000	-7.2
Aroclor-1260-2	6.548	6.448	6.648	439.840	500.000	-12.0
Aroclor-1260-3	6.703	6.602	6.802	387.310	500.000	-22.5
Aroclor-1260-4	7.172	7.072	7.272	381.040	500.000	-23.8
Aroclor-1260-5	7.411	7.311	7.511	378.500	500.000	-24.3
Decachlorobiphenyl	8.813	8.715	8.915	36.350	50.000	-27.3
Tetrachloro-m-xylene	3.801	3.699	3.899	56.810	50.000	13.6