

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD081825

Review By	Abdul	Review On	8/19/2025 9:47:27 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:23 AM
SubDirectory	PD081825	HP Acquire Method	HP Processing Method pd081825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089960.D	18 Aug 2025 10:30		AR\AJ	Ok
2	I.BLK	I.BLK	PD089961.D	18 Aug 2025 10:44		AR\AJ	Ok,M
3	PEM	PEM	PD089962.D	18 Aug 2025 10:57		AR\AJ	Ok
4	RESCHK	RESCHK	PD089963.D	18 Aug 2025 11:11		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD089964.D	18 Aug 2025 11:25		AR\AJ	Ok,M
6	PSTDICC075	PSTDICC075	PD089965.D	18 Aug 2025 11:39		AR\AJ	Ok,M
7	PSTDICC050	PSTDICC050	PD089966.D	18 Aug 2025 11:53		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD089967.D	18 Aug 2025 12:06		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD089968.D	18 Aug 2025 12:20		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD089969.D	18 Aug 2025 13:50	DCB high	AR\AJ	Not Ok
11	PCHLORICC750	PCHLORICC750	PD089970.D	18 Aug 2025 14:06	DCB high	AR\AJ	Not Ok
12	PCHLORICC500	PCHLORICC500	PD089971.D	18 Aug 2025 14:19	Not used	AR\AJ	Not Ok
13	PCHLORICC250	PCHLORICC250	PD089972.D	18 Aug 2025 14:33	DCB high	AR\AJ	Not Ok
14	PCHLORICC050	PCHLORICC050	PD089973.D	18 Aug 2025 14:47	DCB high	AR\AJ	Not Ok
15	PTOXICC1000	PTOXICC1000	PD089974.D	18 Aug 2025 15:00		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD089975.D	18 Aug 2025 15:17		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD089976.D	18 Aug 2025 15:31		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD089977.D	18 Aug 2025 15:44		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD081825

Review By	Abdul	Review On	8/19/2025 9:47:27 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:05:23 AM
SubDirectory	PD081825	HP Acquire Method	HP Processing Method pd081825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD089978.D	18 Aug 2025 15:58		AR\AJ	Ok
20	PSTDICV050	ICVPD081825	PD089979.D	18 Aug 2025 16:12		AR\AJ	Ok
21	PCHLORICV500	ICVPD081825CHLOR	PD089980.D	18 Aug 2025 16:25	Not used	AR\AJ	Not Ok
22	PTOXICV500	ICVPD081825TOX	PD089981.D	18 Aug 2025 16:39		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD081925

Review By	Abdul	Review On	8/20/2025 8:42:02 AM
Supervise By	mohammad	Supervise On	8/22/2025 2:04:20 AM
SubDirectory	PD081925	HP Acquire Method	HP Processing Method pd081825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089982.D	19 Aug 2025 09:36		AR\AJ	Ok
2	I.BLK	I.BLK	PD089983.D	19 Aug 2025 09:50		AR\AJ	Ok
3	PEM	PEM	PD089984.D	19 Aug 2025 10:03		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD089985.D	19 Aug 2025 10:17		AR\AJ	Ok
5	PCHLORICC1000	PCHLORICC1000	PD089986.D	19 Aug 2025 10:30		AR\AJ	Ok
6	PCHLORICC750	PCHLORICC750	PD089987.D	19 Aug 2025 11:10		AR\AJ	Ok
7	PCHLORICC500	PCHLORICC500	PD089988.D	19 Aug 2025 11:24		AR\AJ	Ok
8	PCHLORICC250	PCHLORICC250	PD089989.D	19 Aug 2025 11:37		AR\AJ	Ok
9	PCHLORICC050	PCHLORICC050	PD089990.D	19 Aug 2025 11:51		AR\AJ	Ok,M
10	PCHLORICV500	ICVPD081825CHLOR	PD089991.D	19 Aug 2025 12:28		AR\AJ	Ok
11	I.BLK	I.BLK	PD089992.D	19 Aug 2025 13:29		AR\AJ	Ok
12	PEM	PEM	PD089993.D	19 Aug 2025 13:43		AR\AJ	Ok,M
13	PSTDCCC050	PSTDCCC050	PD089994.D	19 Aug 2025 13:56		AR\AJ	Ok
14	PCHLORCCC500	PCHLORCCC500	PD089995.D	19 Aug 2025 14:48		AR\AJ	Ok
15	PTOXCCC500	PTOXCCC500	PD089996.D	19 Aug 2025 15:04		AR\AJ	Ok,M
16	PB169206BS	PB169206BS	PD089997.D	19 Aug 2025 16:02	CHLOR BS	AR\AJ	Ok
17	PB169206BS	PB169206BS	PD089998.D	19 Aug 2025 16:16	TOX BS	AR\AJ	Ok
18	I.BLK	I.BLK	PD089999.D	19 Aug 2025 16:38		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD081925

Review By	Abdul	Review On	8/20/2025 8:42:02 AM
Supervise By	mohammad	Supervise On	8/22/2025 2:04:20 AM
SubDirectory	PD081925	HP Acquire Method	HP Processing Method pd081825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	PSTDCCC050	PSTDCCC050	PD090000.D	19 Aug 2025 16:52		AR\AJ	Ok,M
20	PCHLORCCC500	PCHLORCCC500	PD090001.D	19 Aug 2025 17:05		AR\AJ	Ok
21	PTOXCCC500	PTOXCCC500	PD090002.D	19 Aug 2025 17:19		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD090225

Review By	Abdul	Review On	9/3/2025 9:59:21 AM
Supervise By		Supervise On	
SubDirectory	PD090225	HP Acquire Method	HP Processing Method pd081825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090104.D	02 Sep 2025 10:38		AR\AJ	Ok
2	I.BLK	I.BLK	PD090105.D	02 Sep 2025 11:05		AR\AJ	Ok
3	PEM	PEM	PD090106.D	02 Sep 2025 11:18		AR\AJ	Ok
4	PSTDCCC050	PSTDCCC050	PD090107.D	02 Sep 2025 11:37		AR\AJ	Ok,NS
5	Q2971-37	GROUNDWATER	PD090108.D	02 Sep 2025 12:54	DCB Low in 2nd column	AR\AJ	Ok
6	PB169488BL	PB169488BL	PD090109.D	02 Sep 2025 13:08		AR\AJ	Ok
7	PB169488BS	PB169488BS	PD090110.D	02 Sep 2025 14:13		AR\AJ	Ok
8	Q2989-01	SOUTH-TP-5	PD090111.D	02 Sep 2025 14:27		AR\AJ	Ok,NS
9	Q2989-01MS	SOUTH-TP-5MS	PD090112.D	02 Sep 2025 14:51		AR\AJ	Ok,NS
10	Q2989-01MSD	SOUTH-TP-5MSD	PD090113.D	02 Sep 2025 15:05		AR\AJ	Ok,NS
11	Q2989-03	SOUTH-TP-6	PD090114.D	02 Sep 2025 15:19		AR\AJ	Ok,NS
12	Q2991-01	309	PD090115.D	02 Sep 2025 15:32		AR\AJ	Ok,NS
13	Q2995-01	RT5427	PD090116.D	02 Sep 2025 15:46		AR\AJ	Ok,NS
14	Q2997-01	EO-01-082925	PD090117.D	02 Sep 2025 15:59		AR\AJ	Ok,NS
15	Q2994-01	PAULDING-WC1	PD090118.D	02 Sep 2025 17:00		AR\AJ	Ok,NS
16	Q2996-01	CONTAINMENT-A	PD090119.D	02 Sep 2025 17:13		AR\AJ	Ok,NS
17	I.BLK	I.BLK	PD090120.D	02 Sep 2025 17:27		AR\AJ	Ok
18	PSTDCCC050	PSTDCCC050	PD090121.D	02 Sep 2025 17:41		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD090225

Review By	Abdul	Review On	9/3/2025 9:59:21 AM
Supervise By		Supervise On	
SubDirectory	PD090225	HP Acquire Method	HP Processing Method pd081825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2996-03	CONTAINMENT-B	PD090122.D	02 Sep 2025 17:57		AR\AJ	Ok,NS
20	Q2996-05	CONTAINMENT-C	PD090123.D	02 Sep 2025 18:11		AR\AJ	Ok,NS
21	Q2998-01	AU-05-082925	PD090124.D	02 Sep 2025 18:25		AR\AJ	Ok,NS
22	Q2996-01	CONTAINMENT-A	PD090125.D	02 Sep 2025 18:38		AR\AJ	Ok,NS
23	I.BLK	I.BLK	PD090126.D	02 Sep 2025 18:52		AR\AJ	Ok
24	PSTDCCC050	PSTDCCC050	PD090127.D	02 Sep 2025 19:06		AR\AJ	Ok

M : Manual Integration