

Surrogate Summary

SDG No.: Q2994

Client: Core Environmental Consultants and Ser

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089961.D	PIBLK-PD089961.D	Decachlorobiphen	1	20	19.3	96		57	171
		Tetrachloro-m-xyl	1	20	17.4	87		61	148
		Decachlorobiphen	2	20	18.3	91		57	171
		Tetrachloro-m-xyl	2	20	17.5	88		61	148
I.BLK-PD090105.D	PIBLK-PD090105.D	Decachlorobiphen	1	20	19.0	95		57	171
		Tetrachloro-m-xyl	1	20	18.4	92		61	148
		Decachlorobiphen	2	20	17.8	89		57	171
		Tetrachloro-m-xyl	2	20	19.9	99		61	148
PB169488BL	PB169488BL	Decachlorobiphen	1	20	16.6	83		20	144
		Tetrachloro-m-xyl	1	20	16.3	81		19	148
		Decachlorobiphen	2	20	16.1	80		20	144
		Tetrachloro-m-xyl	2	20	17.3	86		19	148
PB169488BS	PB169488BS	Decachlorobiphen	1	20	22.1	111		20	144
		Tetrachloro-m-xyl	1	20	20.9	104		19	148
		Decachlorobiphen	2	20	21.4	107		20	144
		Tetrachloro-m-xyl	2	20	21.4	107		19	148
Q2989-01MS	SOUTH-TP-5MS	Decachlorobiphen	1	20	15.5	77		20	144
		Tetrachloro-m-xyl	1	20	16.9	84		19	148
		Decachlorobiphen	2	20	15.1	76		20	144
		Tetrachloro-m-xyl	2	20	16.8	84		19	148
Q2989-01MSD	SOUTH-TP-5MSD	Decachlorobiphen	1	20	15.1	75		20	144
		Tetrachloro-m-xyl	1	20	16.8	84		19	148
		Decachlorobiphen	2	20	15.0	75		20	144
		Tetrachloro-m-xyl	2	20	16.6	83		19	148
Q2994-01	PAULDING-WC1	Decachlorobiphen	1	20	9.31	47		20	144
		Tetrachloro-m-xyl	1	20	9.76	49		19	148
		Decachlorobiphen	2	20	7.12	36		20	144
		Tetrachloro-m-xyl	2	20	11.1	55		19	148
I.BLK-PD090120.D	PIBLK-PD090120.D	Decachlorobiphen	1	20	17.8	89		57	171
		Tetrachloro-m-xyl	1	20	17.7	88		61	148
		Decachlorobiphen	2	20	13.8	69		57	171
		Tetrachloro-m-xyl	2	20	19.3	96		61	148